



## Certificate of Analysis

Product: PSA ELISA (96T)  
Catalogue Number: BXE0851A  
Batch Number: PS-1611-1  
Manufactured Date: 2016/11  
Expiry: 2018/02

Kit Performance:

SAMPLE	QC TEST	RESULTS	SPECIFICATIONS	PASS / FAIL
CONTROL SET 1-1	CONCENTRATION	0.79 NG/ML	0.82 ± 0.36	PASS
CONTROL SET 1-2	CONCENTRATION	3.10 MG/ML	3.12 ± 1.1	PASS
CONTROL SET 1-3	CONCENTRATION	21.3 NG/ML	21.0 ± 6.0	PASS

QC PASS/FAIL: **PASS**

This product has been tested by the Quality Control Department and has met all final QC specifications. This product is released for packing and is suitable for use until the stated expiry date.

This product was produced according to ISO 13485 guidelines.

Products containing Human Source Material have been tested at the donor level for HbsAg and antibodies to HIV1, HIV2 and HCV and found to be non-reactive.

Signed: *Cherith Morris* (QC Manager)

Date: 2016/11