


REF : 3734-3735-3736-3752-3753-3754-3820-3821-3822  
 LOT : 189  
 : 2009, November 19th

**Coulter Counter Models (with III Diff. option; for AcT see separate table)**

**MELET MS9** NOTE : Please refer to next page for Erma Values

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	20.8	2.0
RBC 10 <sup>12</sup> /l	2.40	0.12	4.66	0.15	5.60	0.18
Hb g/dl	6.2	0.4	13.6	0.5	17.7	0.6
Hb mmol/l	3.8	0.2	8.4	0.3	11.0	0.4
Hct %	17.5	1.5	38.7	2.4	51.0	3.0
Hct l/l	0.175	0.015	0.387	0.024	0.510	0.030
MCV fl	73	4	83	4	91	4
MCH pg	25.8	2.4	29.2	2.8	31.6	2.8
MCH fmol	1.60	0.14	1.81	0.11	1.96	0.11
MCHC g/dl	35.4	3.0	35.2	3.0	34.7	3.0
MCHC mmol/l	22.0	1.8	21.8	1.8	21.6	1.8
RDW %	15.6	3.0	15.0	3.0	13.8	3.0
Plts 10 <sup>9</sup> /l	68	15	245	30	515	55
MPV fl	7.8	2.0	7.9	1.5	8.1	1.5
PDW %	17.3	2.5	16.7	2.5	16.9	2.5
LYMP %	59.0	5.0	31.0	4.0	16.5	4.0
MONO %	10.5	4.0	6.7	4.0	4.0	4.0
GRAN %	30.5	5.0	62.3	5.0	79.5	5.0
LYMP #	1.2	0.2	2.4	0.4	3.4	1.0
MONO #	0.2	0.1	0.5	0.5	0.8	0.8
GRAN #	0.6	0.2	4.8	0.6	16.5	1.2

**ABX Roche: Micros; Bayer: Advia 60**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.2	0.4	7.3	0.8	19.3	2.0
RBC 10 <sup>12</sup> /l	2.41	0.12	4.62	0.15	5.59	0.18
Hb g/dl	5.9	0.4	13.6	0.5	17.6	0.6
Hb mmol/l	3.6	0.2	8.4	0.3	10.9	0.4
Hct %	16.3	1.5	37.3	2.4	50.3	3.0
Hct l/l	0.163	0.015	0.373	0.024	0.503	0.030
MCV fl	68	4	81	4	90	4
MCH pg	24.4	2.4	29.4	2.8	31.5	2.8
MCH fmol	1.52	0.14	1.82	0.11	1.96	0.11
MCHC g/dl	36.1	3.0	36.4	3.0	35.0	3.0
MCHC mmol/l	22.4	1.8	22.6	1.8	21.7	1.8
RDW %	14.3	3.0	15.4	3.0	14.2	3.0
Plts 10 <sup>9</sup> /l	74	15	243	30	481	55
MPV fl	7.5	2.0	7.5	1.5	7.6	1.5
LYMP %	51.0	5.0	27.5	4.0	15.0	4.0
MONO %	10.0	4.0	6.6	4.0	3.1	3.0
GRAN %	39.0	4.0	65.9	5.0	81.8	5.0
LYMP #	1.1	0.3	2.0	0.4	2.9	0.8
MONO #	0.2	0.1	0.5	0.3	0.6	0.6
GRAN #	0.8	0.2	4.8	0.6	15.8	1.2

**Abbott Cell-Dyn: 1300, 1500, 1600, 1700, 2000 series**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.3	0.4	7.4	0.8	18.7	2.0
RBC 10 <sup>12</sup> /l	2.43	0.12	4.62	0.15	5.49	0.18
Hb g/dl	6.1	0.4	13.4	0.5	17.5	0.6
Hb mmol/l	3.8	0.2	8.3	0.3	10.9	0.4
Hct %	17.7	1.5	38.8	2.4	50.5	3.0
Hct l/l	0.177	0.015	0.388	0.024	0.505	0.03
MCV fl	73	4	84	4	92	4
MCH pg	25.1	2.4	29.0	2.8	31.9	2.8
MCH fmol	1.56	0.14	1.80	0.11	1.98	0.11
MCHC g/dl	34.4	3.0	34.5	3.0	34.6	3.0
MCHC mmol/l	21.3	1.8	21.4	1.8	21.5	1.8
Plts 10 <sup>9</sup> /l	64	15	240	30	502	55
MPV fl	8.7	2	8.4	1.5	8.3	1.5
LYMP %	59.0	5.0	31.0	4.0	16.5	4.0
MID %	10.5	4.0	6.7	4.0	4.0	3.0
GRAN %	30.5	4.0	62.3	5.0	79.5	5.0
LYMP #	1.4	0.3	2.3	0.4	3.1	0.8
MID #	0.2	0.1	0.5	0.3	0.7	0.6
GRAN #	0.7	0.2	4.6	0.6	14.9	1.2

**Coulter Counter: AcT diff, AcT diff 2**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	20.8	2.0
RBC 10 <sup>12</sup> /l	2.40	0.12	4.66	0.15	5.60	0.18
Hb g/dl	6.2	0.4	13.6	0.5	17.8	0.6
Hb mmol/l	3.8	0.3	8.4	0.3	11.0	0.4
Hct %	17.8	1.5	39.1	2.4	51.5	3.0
Hct l/l	0.178	0.015	0.391	0.024	0.515	0.030
MCV fl	74	4	84	4	92	4
MCH pg	25.8	2.4	29.2	2.8	31.8	2.8
MCH fmol	1.60	0.10	1.81	0.11	1.97	0.12
MCHC g/dl	34.9	3.0	34.7	3.0	34.5	3.0
MCHC mmol/l	21.7	1.8	21.6	1.8	21.4	1.8
RDW %	15.6	3.0	15.0	3.0	13.8	3.0
Plts 10 <sup>9</sup> /l	68	15	235	30	505	55
MPV fl	7.8	2.0	7.9	1.5	8.1	1.5
LYMP %	61.0	5.0	31.0	4.0	14.5	4.0
MONO %	9.0	4.0	5.7	4.0	4.0	4.0
GRAN %	30.0	5.0	63.3	5.0	81.5	5.0
LYMP #	1.3	0.2	2.4	0.6	3.0	1.0
MONO #	0.2	0.1	0.4	0.4	0.8	0.8
GRAN #	0.6	0.2	4.9	0.5	17.0	1.2

**Coulter Counter: AcT 8, AcT 10**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
LYMP %	46.0	6.0	23.7	4.0	10.5	4.0
LYMP #	1.0	0.2	1.8	0.4	2.2	0.9

Please refer to Table AcT Diff for CBC parameters

**Nihon Kohden: Celltac and MEK series (not for 5-part diff)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.4	0.8	19.1	2.0
RBC 10 <sup>12</sup> /l	2.38	0.12	4.62	0.15	5.52	0.18
Hb g/dl	6.1	0.4	13.9	0.5	18.9	0.6
Hb mmol/l	3.8	0.2	8.6	0.3	11.7	0.4
Hct %	17.8	1.5	39.0	2.4	52.8	3.0
Hct l/l	0.178	0.015	0.390	0.024	0.528	0.030
MCV fl	75	4	84	4	96	4
MCH pg	25.6	2.4	30.1	2.8	34.3	2.8
MCH fmol	1.59	0.14	1.87	0.11	2.13	0.11
MCHC g/dl	34.2	3.0	35.6	3.0	35.8	3.0
MCHC mmol/l	21.2	1.8	22.1	1.8	22.2	1.8
RDW %	14.5	3.0	14.5	3.0	13.1	3.0
Plts 10 <sup>9</sup> /l	70	15	250	30	503	55
MPV fl	7.8	2	6.9	1.5	6.6	1.5
LYMP %	62.9	5.0	32.9	4.0	19.9	4.0
MONO %	11.4	4.0	7.3	4.0	3.8	3.0
GRAN %	25.7	4.0	59.8	5.0	76.3	5.0
LYMP #	1.3	0.3	2.4	0.4	3.8	0.8
MONO #	0.2	0.1	0.5	0.3	0.7	0.6
GRAN #	0.5	0.2	4.4	0.6	14.5	1.2

**Danam / Drew Excel18 = MWI BT2100**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.2	0.4	7.8	0.8	21.0	2.0
RBC 10 <sup>12</sup> /l	2.52	0.12	4.58	0.15	5.48	0.18
Hb g/dl	6.2	0.4	13.5	0.5	17.8	0.6
Hb mmol/l	3.9	0.2	8.4	0.3	11.0	0.4
Hct %	18.3	1.5	38.5	2.4	50.6	3.0
Hct l/l	0.183	0.015	0.385	0.024	0.506	0.030
MCV fl	73	4	84	4	92	4
MCH pg	24.8	2.4	29.5	2.8	32.4	2.8
MCH fmol	1.54	0.14	1.83	0.11	2.01	0.11
MCHC g/dl	34.1	3.0	35.1	3.0	35.1	3.0
MCHC mmol/l	21.1	1.8	21.8	1.8	21.8	1.8
RDW	11.8	3.0	11.3	3.0	11.0	3.0
Plts 10 <sup>9</sup> /l	78	15	259	30	501	55
MPV fl	6.5	2	7.1	1.5	7.3	1.5
LYMP %	51.7	5.0	24.9	4.0	12.4	4.0
MONO %	10.9	4.0	6.2	4.0	3.1	3.0
GRAN %	37.4	4.0	68.9	5.0	84.6	5.0
LYMP #	1.1	0.3	1.9	0.4	2.6	0.8
MONO #	0.2	0.1	0.5	0.3	0.6	0.6
GRAN #	0.8	0.2	5.4	0.6	17.7	1.2

LOT : 189

 : 2009, November 19th

Diatron: Abacus, Arcus \*\*

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.4	7.4	0.8	20.4	2.0
RBC 10 <sup>12</sup> /l	2.54	0.12	4.83	0.15	5.76	0.20
Hb g/dl	6.0	0.4	13.3	0.5	17.1	0.6
Hb mmol/l	3.7	0.3	8.3	0.4	10.6	0.4
Hct %	19.8	1.5	42.7	2.4	55.8	3.0
Hct l/l	0.198	0.015	0.427	0.024	0.558	0.030
MCV fl	78	4	88	4	97	4
MCH pg	23.6	2.4	27.6	2.8	29.7	2.8
MCH fmol	1.47	0.14	1.71	0.11	1.84	0.11
MCHC g/dl	30.4	3.0	31.2	3.0	30.7	3.0
MCHC mmol/l	18.8	1.8	19.3	1.8	19.0	1.8
RDW-SD fl	53.4	3.0	58.3	3.0	57.3	3.0
RDW-CV %	17.5	3.0	16.8	3.0	15.0	3.0
Plts 10 <sup>9</sup> /l	65	15	271	30	549	55
MPV fl	7.6	1.5	7.9	1.5	8.3	1.5
PCT %	0.05	0.02	0.21	0.03	0.46	0.04
PCT ml/l	0.49	0.2	2.14	0.3	4.56	0.4
PDW-SD fl	10.4	2.5	10.6	2.5	10.8	3.0
PDW-CV %	38.2	2.5	37.6	2.5	36.5	3.0
LYMP %	56.1	5.0	29.8	4.0	14.0	4.0
MONO %	10.6	4.0	5.9	4.0	4.2	4.0
GRAN %	33.3	4.0	64.2	5.0	81.9	5.0
LYMP #	1.1	0.3	2.2	0.4	2.8	0.8
MONO #	0.2	0.1	0.4	0.3	0.8	0.6
GRAN #	0.6	0.2	4.7	0.6	16.7	1.2

\*\* JT Baker reagents CyMet Abacus + Diluid Abacus = Use Reagent Code 2 in service settings!!!

AVL (Contraves): AVL - 816

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	63.8	5.0	36.9	4.0	19.6	4.0
MONO %	9.6	4.0	7.6	4.0	5.1	3.0
GRAN %	26.6	4.0	55.5	5.0	75.3	5.0
LYMP #	1.3	0.3	2.8	0.4	4.1	0.8
MONO #	0.2	0.1	0.6	0.3	1.1	0.6
GRAN #	0.6	0.2	4.3	0.6	15.7	1.2

AVL (Contraves): AVL - 818

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	71.6	5.0	38.9	4.0	20.7	4.0
MONO %	8.5	4.0	7.1	4.0	4.9	3.0
GRAN %	19.9	4.0	54.0	5.0	74.5	5.0
LYMP #	1.5	0.3	3.0	0.4	4.3	0.8
MONO #	0.2	0.1	0.5	0.2	1.0	0.4
GRAN #	0.4	0.2	4.2	0.6	15.5	1.2

ERMA PCE-170, ERMAX-18, PCE-210

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.9	0.8	21.7	2.0
RBC 10 <sup>12</sup> /l	2.51	0.12	4.67	0.15	5.45	0.18
Hb g/dl	6.0	0.4	13.1	0.5	17.0	0.6
Hb mmol/l	3.7	0.2	8.1	0.3	10.6	0.4
Hct %	19.1	1.5	40.6	2.4	54.0	3.0
Hct l/l	0.191	0.015	0.406	0.024	0.540	0.030
MCV fl	76	4	87	4	99	4
MCH pg	23.9	2.4	28.1	2.8	31.2	2.8
MCH fmol	1.48	0.14	1.74	0.11	1.94	0.11
MCHC g/dl	31.5	3.0	32.2	3.0	31.5	3.0
MCHC mmol/l	19.5	1.8	20.0	1.8	19.6	1.8
RDW %	18.6	3.0	16.5	3.0	15.2	3.0
Plts 10 <sup>9</sup> /l	78	15	259	30	541	55
MPV fl	8.6	2.0	8.3	1.5	8.6	1.5
PDW %	16.1	2.5	15.3	2.5	14.5	2.5
LYMP %	51.0	5.0	32.8	4.0	28.1	4.0
MONO %	4.1	4.1	3.0	3.0	1.0	1.0
GRAN %	44.9	4.0	64.2	5.0	70.9	5.0
LYMP #	1.1	0.3	2.6	0.4	6.1	0.8
MONO #	0.1	0.1	0.2	0.2	0.2	0.2
GRAN #	0.9	0.2	5.1	0.6	15.4	1.2

Orphee MYTHIC 18

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	20.8	2.0
RBC 10 <sup>12</sup> /l	2.40	0.12	4.66	0.15	5.60	0.18
Hb g/dl	6.2	0.4	13.6	0.5	17.7	0.6
Hb mmol/l	3.8	0.2	8.4	0.3	11.0	0.4
Hct %	19.0	1.5	41.0	2.4	53.2	3.0
Hct l/l	0.190	0.015	0.410	0.024	0.532	0.030
MCV fl	79	4	88	4	95	4
MCH pg	25.8	2.4	29.2	2.8	31.6	2.8
MCH fmol	1.60	0.14	1.81	0.11	1.96	0.11
MCHC g/dl	32.7	3.0	33.2	3.0	33.3	3.0
MCHC mmol/l	20.3	1.8	20.6	1.8	20.6	1.8
RDW %	15.6	3.0	15.0	3.0	14.8	3.0
Plts 10 <sup>9</sup> /l	68	15	239	30	503	55
MPV fl	8.5	2.0	8.3	1.5	8.3	1.5
LYMP %	59.3	5.0	29.8	4.0	16.1	4.0
MONO %	5.9	5.9	3.6	3.6	2.0	2.0
GRAN %	34.8	4.0	66.6	5.0	81.9	5.0
LYMP #	1.2	0.3	2.3	0.4	3.3	0.8
MONO #	0.1	0.1	0.3	0.3	0.4	0.4
GRAN #	0.7	0.2	5.1	0.6	17.0	1.2



Mallinckrodt Baker B.V.  
P.O. Box 1 - 7400AA Deventer - Holland  
Tel.: +31 (0)570 687500

The devices as mentioned in this insert comply with the  
In Vitro Diagnostic Medical Devices Directive 98/79/EC.



8°C



LOT : 189

 : 2009, November 19th

**Mindray BC-3200, BC-3000CT (new models)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.5	7.8	1.0	20.8	2.5
RBC 10 <sup>12</sup> /l	2.42	0.18	4.74	0.24	5.75	0.30
Hb g/dl	6.1	0.4	13.6	0.6	18.2	0.8
Hb mmol/l	3.8	0.3	8.4	0.4	11.3	0.5
Hct %	17.9	1.5	40.5	2.0	53.5	2.4
Hct l/l	0.179	0.015	0.405	0.020	0.535	0.024
MCV fl	74.0	5.0	85.5	5.0	93.0	5.0
MCH pg	25.2	2.5	28.7	2.5	31.7	2.5
MCH fmol	1.56	0.16	1.78	0.16	1.96	0.16
MCHC g/dl	34.1	3.0	33.6	3.0	34.0	3.0
MCHC mmol/l	21.1	1.8	20.8	1.8	21.1	1.8
RDW-CV	13.5	3.0	13.0	3.0	12.2	3.0
RDW-SD	35.3	6.0	40.0	6.0	41.0	8.0
Plts 10 <sup>9</sup> /l	74	20	240	40	500	60
MPV fl	7.8	3.0	7.8	3.0	7.8	3.0
LYMP %	60.0	8.0	27.0	7.0	14.0	6.0
MONO %	9.0	7.0	8.5	6.0	5.5	5.5
GRAN %	31.0	7.0	64.5	7.0	80.5	8.0
LYMP #	1.1	0.2	2.1	0.6	2.9	1.3
MONO #	0.2	0.2	0.7	0.6	1.1	1.1
GRAN #	0.6	0.2	5.0	0.6	16.7	1.8

**Mindray BC-2800, BC-2600 (software version lower than 2.0)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.5	7.7	1.0	21.2	2.5
RBC 10 <sup>12</sup> /l	2.33	0.18	4.64	0.24	5.58	0.30
Hb g/dl	6.4	0.4	14.1	0.6	18.5	0.8
Hb mmol/l	4.0	0.3	8.8	0.4	11.5	0.5
Hct %	17.8	1.5	39.9	2.0	52.2	2.4
Hct l/l	0.178	0.015	0.399	0.020	0.522	0.024
MCV fl	76.5	5.0	86.0	5.0	93.5	5.0
MCH pg	27.5	2.5	30.4	2.5	33.2	2.5
MCH fmol	1.70	0.16	1.89	0.16	2.06	0.16
MCHC g/dl	35.9	3.0	35.3	3.0	35.5	3.0
MCHC mmol/l	22.3	1.8	21.9	1.8	22.0	1.8
RDW-CV	14.2	3.0	14.2	3.0	13.5	3.0
RDW-SD	37.5	6.0	42.0	6.0	43.5	8.0
Plts 10 <sup>9</sup> /l	65	20	240	40	540	60
MPV fl	9.4	3.0	9.4	3.0	9.3	3.0
PCT %	0.06	0.05	0.23	0.10	0.50	0.20
PCT ml/l	0.61	0.50	2.26	1.00	5.02	2.00
PDW fl	16.3	3.0	16.1	3.0	16.0	3.0
LYMP %	61.0	8.0	30.0	7.0	14.7	6.0
MONO %	11.0	7.0	11.5	6.0	9.6	5.0
GRAN %	28.0	7.0	58.5	7.0	75.7	8.0
LYMP #	1.1	0.2	2.3	0.6	3.1	1.3
MONO #	0.2	0.2	0.9	0.5	2.0	1.1
GRAN #	0.5	0.2	4.5	0.6	16.0	1.7

**Mindray BC-2800, BC-2600 (software version 2.0 or higher)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	7.8	1.0	20.8	2.5
RBC 10 <sup>12</sup> /l	2.38	0.18	4.70	0.24	5.60	0.30
Hb g/dl	6.3	0.4	13.8	0.6	18.3	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.4	0.5
Hct %	17.4	1.5	39.2	2.0	51.8	2.4
Hct l/l	0.174	0.015	0.392	0.020	0.518	0.024
MCV fl	73.0	5.0	83.5	5.0	92.5	5.0
MCH pg	26.5	2.5	29.4	2.5	32.7	2.5
MCH fmol	1.64	0.16	1.82	0.16	2.03	0.16
MCHC g/dl	36.3	3.0	35.2	3.0	35.3	3.0
MCHC mmol/l	22.5	1.8	21.8	1.8	21.9	1.8
RDW-CV	16.0	3.0	15.5	3.0	14.5	3.0
RDW-SD	40.0	6.0	44.5	6.0	45.0	8.0
Plts 10 <sup>9</sup> /l	70	20	245	40	500	60
MPV fl	9.7	3.0	9.6	3.0	9.7	3.0
PCT %	0.07	0.05	0.24	0.10	0.49	0.20
PCT ml/l	0.68	0.50	2.35	1.00	4.85	2.00
PDW fl	14.7	3.0	14.8	3.0	14.8	3.0
LYMP %	59.5	9.0	29.0	8.0	15.0	6.0
MONO %	9.5	8.0	6.5	6.0	4.5	4.5
GRAN %	31.0	8.0	64.5	9.0	80.5	8.0
LYMP #	1.2	0.2	2.3	0.7	3.1	1.3
MONO #	0.2	0.2	0.5	0.5	0.9	0.9
GRAN #	0.6	0.2	5.0	0.8	16.7	1.8

**Mindray BC-3000 (new models s/n A75xxx) , BC-2900, BC-1800**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	7.8	1.0	20.4	2.5
RBC 10 <sup>12</sup> /l	2.49	0.18	4.85	0.24	5.75	0.30
Hb g/dl	6.3	0.4	13.8	0.6	18.2	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.3	0.5
Hct %	18.9	1.5	42.0	2.0	54.1	2.4
Hct l/l	0.189	0.015	0.420	0.020	0.541	0.024
MCV fl	76.0	5.0	86.5	5.0	94.0	5.0
MCH pg	25.3	2.5	28.5	2.5	31.7	2.5
MCH fmol	1.57	0.16	1.77	0.16	1.96	0.16
MCHC g/dl	33.3	3.0	32.9	3.0	33.7	3.0
MCHC mmol/l	20.7	1.8	20.4	1.8	20.9	1.8
RDW-CV	13.5	3.0	13.5	3.0	13.0	3.0
RDW-SD	37.0	6.0	43.0	6.0	43.5	8.0
Plts 10 <sup>9</sup> /l	77	20	246	40	500	60
MPV fl	8.8	3.0	8.7	3.0	8.8	3.0
PCT %	0.07	0.05	0.21	0.10	0.44	0.20
PCT ml/l	0.68	0.50	2.14	1.00	4.40	2.00
PDW fl	16.2	3.0	15.7	3.0	15.8	3.0
LYMP %	57.0	10.0	28.0	8.0	14.5	7.0
MONO %	11.5	9.0	8.5	6.0	5.5	5.5
GRAN %	31.5	8.0	63.5	9.0	80.0	8.0
LYMP #	1.1	0.3	2.2	0.7	3.0	1.5
MONO #	0.2	0.2	0.7	0.6	1.1	1.1
GRAN #	0.6	0.2	5.0	0.8	16.3	1.7

**Mindray: BC-3000 with Cymet Mindray CN-free (old models s/n A59xxx)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.5	7.7	1.0	19.6	2.5
RBC 10 <sup>12</sup> /l	2.25	0.18	4.49	0.24	5.54	0.30
Hb g/dl **	6.4	0.4	14.5	0.6	19.1	0.8
Hb mmol/l **	4.0	0.3	9.0	0.4	11.9	0.5
Hct %	17.8	1.5	40.0	2.0	53.2	2.4
Hct l/l	0.178	0.015	0.400	0.020	0.532	0.024
MCV fl	79.0	5.0	89.0	5.0	96.0	5.0
MCH pg **	28.4	2.5	32.3	2.5	34.5	2.5
MCH fmol **	1.77	0.16	2.00	0.16	2.14	0.16
MCHC g/dl **	36.0	3.0	36.3	3.0	35.9	3.0
MCHC mmol/l **	22.3	1.8	22.5	1.8	22.3	1.8
RDW-CV	14.4	3.0	14.3	3.0	13.6	3.0
RDW-SD	40.6	6.0	45.5	6.0	46.6	6.0
Plts 10 <sup>9</sup> /l	75	20	227	40	471	60
MPV fl	8.5	3.0	8.7	3.0	8.6	3.0
LYMP %	59.0	8.0	28.0	7.0	14.0	6.0
MONO %	10.5	7.0	6.7	6.0	4.0	4.0
GRAN %	30.5	7.0	65.3	7.0	82.0	8.0
LYMP #	1.2	0.2	2.2	0.6	2.7	1.3
MONO #	0.2	0.2	0.5	0.5	0.8	0.8
GRAN #	0.6	0.2	5.0	0.6	16.1	1.8

**\*\* Mindray: BC-3000 with Cymet Mindray (with CN) (old series s/n A59xxx)**

Hb g/dl	6.1	0.4	13.5	0.6	17.8	0.8
Hb mmol/l	3.8	0.3	8.4	0.4	11.0	0.5
MCH pg	27.1	2.5	30.1	2.5	32.1	2.5
MCH fmol	1.68	0.16	1.87	0.16	1.99	0.16
MCHC g/dl	34.3	3.0	33.8	3.0	33.5	3.0
MCHC mmol/l	21.3	1.8	21.0	1.8	20.8	1.8

**Mindray BC-2300, BC-2100**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.5	7.6	1.0	20.9	2.5
RBC 10 <sup>12</sup> /l	2.35	0.18	4.55	0.24	5.45	0.30
Hb g/dl	6.3	0.4	13.8	0.6	18.2	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.3	0.5
Hct %	16.7	1.5	37.8	2.0	49.6	2.4
Hct l/l	0.167	0.015	0.378	0.020	0.496	0.024
MCV fl	71.0	5.0	83.0	5.0	91.0	5.0
MCH pg	26.8	2.5	30.3	2.5	33.4	2.5
MCH fmol	1.66	0.16	1.88	0.16	2.07	0.16
MCHC g/dl	37.8	3.0	36.5	3.0	36.7	3.0
MCHC mmol/l	23.4	1.8	22.7	1.8	22.8	1.8
RDW-CV	13.3	3.0	13.4	3.0	12.5	3.0
RDW-SD	27.5	6.0	30.5	6.0	31.0	8.0
Plts 10 <sup>9</sup> /l	65	20	230	40	490	60
MPV fl	11.4	3.0	11.0	3.0	11.1	3.0
PCT %	0.07	0.05	0.25	0.10	0.54	0.20
PCT ml/l	0.74	0.50	2.53	1.00	5.44	2.00
PDW fl	14.9	3.0	14.7	3.0	14.8	3.0
LYMP %	56.0	9.0	28.2	8.0	15.0	6.0
MONO %	11.0	8.0	9.8	8.0	6.0	5.0
GRAN %	33.0	8.0	62.0	9.0	79.0	8.0
LYMP #	1.1	0.2	2.1	0.7	3.1	1.3
MONO #	0.2	0.2	0.7	0.7	1.3	1.3
GRAN #	0.6	0.2	4.7	0.8	16.5	1.8