

REF : 3433-3434-3435-3466-3467-3468-3502-3503-3504

LOT : 205

: 2013-11-19

Coulter Counter Models (with III Diff. option; for AcT see separate table)

MELET MS9

NOTE : Please refer to next page for Erma Values

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.5	2.0
RBC 10 ¹² /l	2.40	0.15	4.60	0.20	5.65	0.24
Hb g/dl	5.9	0.4	13.5	0.5	18.1	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.2	0.5
Hct %	17.6	1.5	39.1	2.4	52.3	3.0
Hct l/l	0.176	0.015	0.391	0.024	0.523	0.030
MCV fl	73.5	4.0	85.0	4.0	92.5	4.0
MCH pg	24.6	2.4	29.3	2.8	32.0	2.8
MCH fmol	1.53	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	33.4	3.0	34.5	3.0	34.6	3.0
MCHC mmol/l	20.8	2.0	21.4	2.0	21.5	2.0
RDW %	16.6	3.0	16.0	3.0	14.5	3.0
Plts 10 ⁹ /l	64	20	252	35	512	60
MPV fl	7.7	2.0	8.1	2.0	8.1	2.0
PDW %	18.3	3.0	16.9	3.0	16.6	3.0
LYMP %	58.5	6.0	31.4	4.0	15.3	4.0
MONO %	10.3	4.0	6.6	4.0	4.4	3.0
GRAN %	31.2	5.0	62.0	5.0	80.3	5.0
LYMP #	1.2	0.2	2.4	0.4	3.1	0.9
MONO #	0.2	0.2	0.5	0.4	0.9	0.7
GRAN #	0.7	0.2	4.8	0.4	16.5	1.1

ABX Roche: Micros 60, Micros ES 60; Bayer: Advia 60

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.3	0.8	19.0	2.0
RBC 10 ¹² /l	2.41	0.15	4.56	0.20	5.64	0.24
Hb g/dl	5.6	0.4	13.5	0.5	18.0	0.6
Hb mmol/l	3.5	0.3	8.4	0.4	11.2	0.5
Hct %	16.4	1.5	37.7	2.4	51.6	3.0
Hct l/l	0.164	0.015	0.377	0.024	0.516	0.030
MCV fl	68.2	4.0	82.6	4.0	91.5	4.0
MCH pg	23.2	2.4	29.6	2.8	31.9	2.8
MCH fmol	1.44	0.16	1.83	0.17	1.98	0.19
MCHC g/dl	34.1	3.0	35.8	3.0	34.9	3.0
MCHC mmol/l	21.2	2.0	22.2	2.0	21.7	2.0
RDW %	15.2	3.0	16.4	3.0	14.9	3.0
Plts 10 ⁹ /l	70	20	250	35	479	60
MPV fl	7.4	2.0	7.7	2.0	7.6	2.0
LYMP %	50.6	6.0	27.9	4.0	13.9	4.0
MONO %	9.8	4.0	6.5	4.0	3.5	3.0
GRAN %	39.6	5.0	65.6	5.0	82.6	5.0
LYMP #	1.1	0.2	2.0	0.4	2.6	0.9
MONO #	0.2	0.2	0.5	0.5	0.7	0.7
GRAN #	0.9	0.2	4.8	0.4	15.7	1.1

Abbott Cell-Dyn: 1300, 1500, 1600, 1700, 1800 series

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.3	0.4	7.4	0.8	18.4	2.0
RBC 10 ¹² /l	2.43	0.15	4.56	0.20	5.54	0.24
Hb g/dl	5.8	0.4	13.3	0.5	17.9	0.6
Hb mmol/l	3.6	0.3	8.3	0.4	11.1	0.5
Hct %	17.9	1.5	39.2	2.4	51.8	3.0
Hct l/l	0.179	0.015	0.392	0.024	0.518	0.030
MCV fl	73.5	4.0	86.0	4.0	93.5	4.0
MCH pg	23.9	2.4	29.2	2.8	32.3	2.8
MCH fmol	1.48	0.16	1.81	0.17	2.01	0.19
MCHC g/dl	32.5	3.0	33.9	3.0	34.6	3.0
MCHC mmol/l	20.2	2.0	21.0	2.0	21.4	2.0
Plts 10 ⁹ /l	60	20	247	35	499	60
MPV fl	8.6	2.0	8.6	2.0	8.3	2.0
LYMP %	58.5	6.0	31.4	4.0	15.3	4.0
MID %	10.3	4.0	6.6	4.0	4.4	3.0
GRAN %	31.2	5.0	62.0	5.0	80.3	5.0
LYMP #	1.3	0.2	2.3	0.4	2.8	0.9
MID #	0.2	0.2	0.5	0.4	0.8	0.7
GRAN #	0.7	0.2	4.6	0.4	14.8	1.1

Coulter Counter: AcT diff, AcT diff 2

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.4	2.0
RBC 10 ¹² /l	2.40	0.15	4.60	0.20	5.65	0.24
Hb g/dl	5.9	0.4	13.5	0.5	18.1	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.2	0.5
Hct %	17.6	1.5	39.6	2.4	52.8	3.0
Hct l/l	0.176	0.015	0.396	0.024	0.528	0.030
MCV fl	73.5	4.0	86.0	4.0	93.5	4.0
MCH pg	24.6	2.4	29.3	2.8	32.0	2.8
MCH fmol	1.52	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	33.4	3.0	34.1	3.0	34.3	3.0
MCHC mmol/l	20.7	2.0	21.2	2.0	21.2	2.0
RDW %	16.6	3.0	16.0	3.0	14.5	3.0
Plts 10 ⁹ /l	64	20	250	35	510	60
MPV fl	7.7	2.0	8.1	2.0	8.1	2.0
LYMP %	60.2	6.0	31.0	4.0	14.0	4.0
MONO %	7.5	4.0	5.0	4.0	3.5	3.0
GRAN %	32.3	5.0	64.0	5.0	82.5	5.0
LYMP #	1.3	0.2	2.4	0.4	2.9	0.9
MONO #	0.2	0.2	0.4	0.4	0.7	0.7
GRAN #	0.7	0.2	4.9	0.4	16.8	1.1

Coulter Counter: AcT 8, AcT 10

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
LYMP %	46.7	6.0	23.0	4.0	9.5	4.0
LYMP #	1.0	0.2	1.8	0.4	1.9	0.8

Please refer to Table AcT Diff for CBC parameters

Nihon Kohden: Celltac and MEK series (not for 5-part diff)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.5	2.0
RBC 10 ¹² /l	2.38	0.15	4.60	0.20	5.56	0.24
Hb g/dl	6.0	0.4	13.8	0.5	18.8	0.6
Hb mmol/l	3.7	0.3	8.6	0.4	11.7	0.5
Hct %	17.5	1.5	39.8	2.4	54.2	3.0
Hct l/l	0.175	0.015	0.398	0.024	0.542	0.030
MCV fl	73.5	4.0	86.4	4.0	97.4	4.0
MCH pg	25.3	2.4	30.0	2.8	33.8	2.8
MCH fmol	1.57	0.16	1.86	0.17	2.10	0.19
MCHC g/dl	34.4	3.0	34.7	3.0	34.7	3.0
MCHC mmol/l	21.4	2.0	21.5	2.0	21.6	2.0
RDW %	15.5	3.0	15.4	3.0	13.7	3.0
Plts 10 ⁹ /l	65	20	257	35	532	60
MPV fl	7.7	2.0	7.1	2.0	6.6	2.0
LYMP %	62.3	8.0	32.7	4.0	17.9	4.0
MONO %	12.9	7.0	7.1	4.0	4.2	3.0
GRAN %	24.8	8.0	60.2	5.0	77.9	5.0
LYMP #	1.3	0.2	2.5	0.4	3.7	0.9
MONO #	0.3	0.3	0.6	0.4	0.9	0.7
GRAN #	0.5	0.4	4.6	0.4	16.0	1.1

Danam/Drew Excel18=MWI BT2100, D3; Orphee Mythic 18; Abbott Emerald 18

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.5	2.0
RBC 10 ¹² /l	2.40	0.15	4.60	0.20	5.65	0.24
Hb g/dl	5.7	0.4	13.5	0.5	18.1	0.6
Hb mmol/l	3.5	0.3	8.4	0.4	11.2	0.5
Hct %	17.1	1.5	39.6	2.4	53.0	3.0
Hct l/l	0.171	0.015	0.396	0.024	0.530	0.030
MCV fl	71.3	4.0	86.0	4.0	93.9	4.0
MCH pg	23.7	2.4	29.3	2.8	32.0	2.8
MCH fmol	1.47	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	33.3	3.0	34.1	3.0	34.1	3.0
MCHC mmol/l	20.7	2.0	21.2	2.0	21.2	2.0
RDW %	14.7	3.0	13.5	3.0	12.2	3.0
Plts 10 ⁹ /l	64	20	235	35	475	60
MPV fl	7.7	2.0	8.1	2.0	8.1	2.0
LYMP %	58.8	6.0	30.2	4.0	14.9	4.0
MONO %	5.7	4.0	3.5	3.5	2.4	2.4
GRAN %	35.5	5.0	66.3	5.0	82.7	5.0
LYMP #	1.2	0.2	2.3	0.4	3.1	0.9
MONO #	0.1	0.1	0.3	0.3	0.5	0.5
GRAN #	0.7	0.2	5.1	0.4	17.0	1.1

LOT : 205
⏳ : 2013-11-19

Diatron[®]: Abacus, Abacus 3, Abacus Junior, Arcus; Erba: Elite 3

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.4	7.5	0.8	20.5	2.0
RBC 10 ¹² /l	2.45	0.15	4.74	0.20	5.75	0.24
Hb g/dl	6.1	0.4	13.6	0.5	18.6	0.6
Hb mmol/l	3.8	0.3	8.5	0.4	11.6	0.5
Hct %	18.7	1.5	42.7	2.4	56.6	3.0
Hct l/l	0.187	0.015	0.427	0.024	0.566	0.030
MCV fl	76.5	4.0	90.0	4.0	98.5	4.0
MCH pg	25.1	2.4	28.7	2.8	32.4	2.8
MCH fmol	1.56	0.16	1.78	0.17	2.01	0.19
MCHC g/dl	32.8	3.0	31.9	3.0	32.9	3.0
MCHC mmol/l	20.3	2.0	19.8	2.0	20.4	2.0
RDW-SD fl	56.7	5.0	62.3	5.0	62.4	5.0
RDW-CV %	16.3	3.0	15.3	3.0	14.0	3.0
Plts 10 ⁹ /l	59	20	252	35	512	60
MPV fl	7.5	2.0	8.1	2.0	7.8	2.0
PCT %	0.04	0.02	0.20	0.06	0.40	0.10
PCT ml/l	0.44	0.2	2.04	0.3	3.97	0.4
PDW-SD fl	10.1	3.0	11.0	3.0	10.5	3.0
PDW-CV %	37.6	3.0	38.0	3.0	38.0	3.0
LYMP %	54.5	6.0	29.4	4.0	12.2	4.0
MONO %	8.3	6.0	6.1	4.0	5.7	3.0
GRAN %	37.2	6.0	64.5	5.0	82.2	6.0
LYMP #	1.1	0.2	2.2	0.4	2.5	0.9
MONO #	0.2	0.2	0.5	0.5	1.2	0.7
GRAN #	0.7	0.2	4.9	0.4	16.8	1.1

** JT Baker reagents CyMet Abacus+Diluid Abacus=Use Reagent Code 2 in service settings!

AVL (Contraves): AVL - 816

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	63.2	6.0	37.4	4.0	18.2	4.0
MONO %	9.5	4.0	7.5	4.0	5.6	3.0
GRAN %	27.3	5.0	55.1	5.0	76.2	5.0
LYMP #	1.3	0.2	2.9	0.4	3.7	0.9
MONO #	0.2	0.2	0.6	0.6	1.2	0.7
GRAN #	0.6	0.2	4.2	0.4	15.6	1.1

AVL (Contraves): AVL - 818

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	71.0	6.0	39.4	4.0	19.2	4.0
MONO %	8.3	4.0	7.0	4.0	5.4	3.0
GRAN %	20.7	5.0	53.6	5.0	75.5	5.0
LYMP #	1.5	0.2	3.0	0.4	3.9	0.9
MONO #	0.2	0.2	0.5	0.5	1.1	0.7
GRAN #	0.4	0.2	4.1	0.4	15.5	1.1

Inmesco HM1060; Rayto RT-7600; HTI Micros CC 20 Plus (all QC-Mode)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.4	2.0
RBC 10 ¹² /l	2.40	0.15	4.50	0.20	5.45	0.24
Hb g/dl	5.9	0.4	14.0	0.5	18.6	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.5	0.5
Hct %	17.6	1.5	39.6	2.4	52.8	3.0
Hct l/l	0.176	0.015	0.396	0.024	0.528	0.030
MCV fl	73.5	4.0	86.0	4.0	93.5	4.0
MCH pg	24.6	2.4	31.1	2.8	34.1	2.8
MCH fmol	1.54	0.2	1.87	0.2	2.12	0.2
MCHC g/dl	33.5	3.0	35.4	3.0	35.2	3.0
MCHC mmol/l	21.0	2.0	21.2	2.0	21.9	2.0
RDW-CV %	19.5	3.0	16.0	3.0	12.4	3.0
RDW-SD fl	50.3	5.0	47.9	5.0	39.3	5.0
Plts 10 ⁹ /l	64	20	240	35	480	60
MPV fl	9.6	2.0	10.4	2.0	10.0	2.0
PCT %	0.06	0.02	0.25	0.06	0.48	0.10
PCT ml/l	0.61	0.20	2.50	0.30	4.80	0.40
PDW %	12.1	3.0	18.1	3.0	18.0	3.0
LYMP %	43.4	8.0	18.7	5.0	9.1	4.0
MONO %	23.0	6.0	13.7	5.0	5.9	5.9
GRAN %	33.6	6.0	67.6	8.0	85.0	8.0
LYMP #	0.9	0.2	1.4	0.4	1.9	0.9
MONO #	0.5	0.5	1.1	0.4	1.2	0.7
GRAN #	0.7	0.2	5.2	0.8	17.3	1.1

Avantor BeneSphera[™] 3 (QC-Mode)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.7	0.8	20.4	2.0
RBC 10 ¹² /l	2.40	0.15	4.50	0.20	5.45	0.24
Hb g/dl	5.9	0.4	14.0	0.5	18.6	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.5	0.5
Hct %	17.6	1.5	39.6	2.4	52.8	3.0
Hct l/l	0.176	0.015	0.396	0.024	0.528	0.030
MCV fl	73.5	4.0	86.0	4.0	93.5	4.0
MCH pg	24.6	2.4	31.1	2.8	34.1	2.8
MCH fmol	1.54	0.2	1.87	0.2	2.12	0.2
MCHC g/dl	33.5	3.0	35.4	3.0	35.2	3.0
MCHC mmol/l	21.0	2.0	21.2	2.0	21.9	2.0
RDW-CV %	19.5	3.0	16.0	3.0	12.4	3.0
RDW-SD fl	50.3	5.0	47.9	5.0	39.3	5.0
Plts 10 ⁹ /l	64	20	240	35	480	60
MPV fl	9.6	2.0	10.4	2.0	10.0	2.0
PCT %	0.06	0.02	0.25	0.06	0.48	0.10
PCT ml/l	0.61	0.20	2.50	0.30	4.80	0.40
PDW %	12.1	3.0	18.1	3.0	18.0	3.0
LYMP %	43.4	8.0	18.7	5.0	9.1	4.0
MONO %	23.0	6.0	13.7	5.0	5.9	5.9
GRAN %	33.6	6.0	67.6	8.0	85.0	8.0
LYMP #	0.9	0.2	1.4	0.4	1.9	0.9
MONO #	0.5	0.5	1.1	0.4	1.2	0.7
GRAN #	0.7	0.2	5.2	0.8	17.3	1.1

ERMA PCE-170, ERMAX-18, PCE-210


Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.4	7.9	0.8	21.4	2.0
RBC 10 ¹² /l	2.51	0.15	4.61	0.20	5.50	0.24
Hb g/dl	5.7	0.4	13.0	0.5	17.4	0.6
Hb mmol/l	3.5	0.3	8.1	0.4	10.8	0.5
Hct %	19.2	1.5	41.0	2.4	55.3	3.0
Hct l/l	0.192	0.015	0.410	0.024	0.553	0.030
MCV fl	76.5	4.0	89.0	4.0	100.5	4.0
MCH pg	22.7	2.4	28.2	2.8	31.6	2.8
MCH fmol	1.41	0.16	1.75	0.17	1.96	0.19
MCHC g/dl	29.7	3.0	31.7	3.0	31.5	3.0
MCHC mmol/l	18.4	2.0	19.7	2.0	19.5	2.0
RDW %	19.6	3.0	17.5	3.0	15.9	3.0
Plts 10 ⁹ /l	74	20	266	35	538	60
MPV fl	8.5	2.0	8.5	2.0	8.6	2.0
PDW %	17.1	3.0	15.5	3.0	14.2	3.0
LYMP %	50.5	6.0	33.2	4.0	26.9	4.0
MONO %	3.9	3.9	2.9	2.9	1.4	1.4
GRAN %	45.6	5.0	63.9	5.0	71.7	5.0
LYMP #	1.1	0.2	2.6	0.4	5.8	0.9
MONO #	0.1	0.1	0.2	0.2	0.3	0.3
GRAN #	1.0	0.2	5.0	0.4	15.3	1.1

Orphee MYTHIC 18

Please refer now to Danam/Drew table

Avantor Performance Materials B.V.
P.O. Box 1 - 7400AA Deventer - The Netherlands
Tel.: +31 (0)570 687500

The devices as mentioned in this insert comply with the
In Vitro Diagnostic Medical Devices Directive 98/79/EC.



2°C

8°C

LOT : 205
: 2013-11-19

Mindray BC-3200, BC-3000CT

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	8.0	1.0	21.0	2.5
RBC 10 ¹² /l	2.43	0.18	4.70	0.24	5.85	0.30
Hb g/dl	6.0	0.4	13.8	0.6	19.0	0.8
Hb mmol/l	3.7	0.3	8.6	0.4	11.8	0.5
Hct %	18.0	1.5	40.9	2.0	56.2	2.4
Hct l/l	0.180	0.015	0.409	0.020	0.562	0.024
MCV fl	74.0	5.0	87.0	5.0	96.0	5.0
MCH pg	24.7	2.5	29.4	2.5	32.5	2.5
MCH fmol	1.53	0.16	1.82	0.16	2.02	0.16
MCHC g/dl	33.4	3.0	33.7	3.0	33.8	3.0
MCHC mmol/l	20.7	1.9	20.9	1.9	21.0	1.9
RDW-CV	14.5	3.0	14.0	3.0	13.0	3.0
RDW-SD	35.5	6.0	39.0	6.0	41.0	8.0
Plts 10 ⁹ /l	70	20	250	40	497	60
MPV fl	8.0	3.0	8.0	3.0	7.6	3.0
LYMP %	58.0	12.0	28.5	8.0	12.5	6.0
MONO %	9.0	8.0	6.5	6.0	6.0	5.0
GRAN %	33.0	9.0	65.0	8.0	81.5	8.0
LYMP #	1.2	0.3	2.3	0.7	2.6	1.3
MONO #	0.2	0.2	0.5	0.5	1.3	1.1
GRAN #	0.7	0.2	5.2	0.7	17.1	1.7

Mindray BC-2800, BC-2600 (software version 2.0 or higher)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.7	1.0	20.9	2.5
RBC 10 ¹² /l	2.38	0.18	4.69	0.24	5.76	0.30
Hb g/dl	6.1	0.4	14.0	0.6	19.2	0.8
Hb mmol/l	3.8	0.3	8.7	0.4	11.9	0.5
Hct %	17.6	1.5	40.6	2.0	54.7	2.4
Hct l/l	0.176	0.015	0.406	0.020	0.547	0.024
MCV fl	74.0	5.0	86.5	5.0	95.0	5.0
MCH pg	25.6	2.5	29.9	2.5	33.3	2.5
MCH fmol	1.59	0.16	1.85	0.16	2.07	0.16
MCHC g/dl	34.6	3.0	34.5	3.0	35.1	3.0
MCHC mmol/l	21.5	1.9	21.4	1.9	21.8	1.9
RDW-CV	16.0	3.0	15.5	3.0	14.5	3.0
RDW-SD	40.5	6.0	45.0	6.0	45.0	8.0
Plts 10 ⁹ /l	58	25	235	40	490	60
MPV fl	9.5	3.0	9.5	3.0	9.4	3.0
PCT %	0.05	0.05	0.22	0.10	0.45	0.20
PCT ml/l	0.50	0.50	2.20	1.00	4.50	2.00
PDW fl	14.7	3.0	14.9	3.0	14.7	3.0
LYMP %	61.5	10.0	31.5	8.0	14.5	7.0
MONO %	6.5	6.0	5.5	5.0	5.0	4.0
GRAN %	32.0	9.0	63.0	9.0	80.5	8.0
LYMP #	1.2	0.2	2.4	0.7	3.0	1.5
MONO #	0.1	0.1	0.4	0.4	1.0	1.0
GRAN #	0.6	0.2	4.9	0.7	16.8	1.7

Mindray BC-2300, BC-2100

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	1.9	0.5	7.6	1.0	20.9	2.5
RBC 10 ¹² /l	2.42	0.18	4.66	0.24	5.67	0.30
Hb g/dl	6.3	0.4	14.1	0.6	18.6	0.8
Hb mmol/l	3.9	0.3	8.8	0.4	11.5	0.5
Hct %	17.3	1.5	39.5	2.0	52.8	2.4
Hct l/l	0.173	0.015	0.395	0.020	0.528	0.024
MCV fl	71.6	5.0	84.7	5.0	93.1	5.0
MCH pg	26.0	2.5	30.3	2.5	32.8	2.5
MCH fmol	1.62	0.16	1.88	0.16	2.04	0.16
MCHC g/dl	36.4	3.0	35.7	3.0	35.2	3.0
MCHC mmol/l	22.6	1.9	22.2	1.9	21.9	1.9
RDW-CV	15.1	3.0	14.3	3.0	13.1	3.0
RDW-SD	34.3	8.0	38.6	8.0	38.5	8.0
Plts 10 ⁹ /l	54	20	242	40	501	60
MPV fl	10.0	3.0	10.1	3.0	9.9	3.0
PCT %	0.05	0.05	0.24	0.10	0.50	0.20
PCT ml/l	0.50	0.50	2.40	1.00	5.00	2.00
PDW fl	15.1	3.0	15.1	3.0	15.0	3.0
LYMP %	60.0	10.0	31.4	8.0	14.1	7.0
MONO %	8.2	6.0	6.9	5.0	6.6	4.0
GRAN %	31.8	9.0	61.7	9.0	79.3	8.0
LYMP #	1.1	0.2	2.4	0.7	2.9	1.5
MONO #	0.2	0.2	0.5	0.4	1.4	0.9
GRAN #	0.6	0.2	4.7	0.7	16.6	1.7

Mindray BC-3000 Plus (new models s/n A75xxx) , BC-2900, BC-1800

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.9	1.0	20.7	2.5
RBC 10 ¹² /l	2.48	0.18	4.76	0.24	5.85	0.30
Hb g/dl	6.0	0.4	13.8	0.6	19.0	0.8
Hb mmol/l	3.7	0.3	8.6	0.4	11.8	0.5
Hct %	18.4	1.5	41.4	2.0	55.9	2.4
Hct l/l	0.184	0.015	0.414	0.020	0.559	0.024
MCV fl	74.0	5.0	87.0	5.0	95.5	5.0
MCH pg	24.2	2.5	29.0	2.5	32.5	2.5
MCH fmol	1.50	0.16	1.80	0.16	2.02	0.16
MCHC g/dl	32.7	3.0	33.3	3.0	34.0	3.0
MCHC mmol/l	20.2	1.9	20.7	1.9	21.1	1.9
RDW-CV	16.0	3.0	15.0	3.0	14.0	3.0
RDW-SD	39.0	6.0	43.0	6.0	44.0	8.0
Plts 10 ⁹ /l	70	20	258	40	510	60
MPV fl	8.9	3.0	8.7	3.0	8.6	3.0
PCT %	0.06	0.05	0.22	0.10	0.43	0.20
PCT ml/l	0.60	0.50	2.20	1.00	4.30	2.00
PDW fl	16.2	3.0	15.8	3.0	15.7	3.0
LYMP %	58.0	12.0	31.0	8.0	13.5	7.0
MONO %	9.0	8.0	6.0	5.0	5.0	4.0
GRAN %	33.0	9.0	63.0	8.0	81.5	8.0
LYMP #	1.2	0.3	2.4	0.7	2.8	1.5
MONO #	0.2	0.2	0.5	0.5	1.0	1.0
GRAN #	0.7	0.2	5.0	0.7	16.9	1.7

Mindray BC-3000 with CyMet Mindray CN-free (old models s/n A59xxx)


Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.5	7.7	1.0	19.3	2.5
RBC 10 ¹² /l	2.25	0.18	4.43	0.24	5.59	0.30
Hb g/dl **	6.1	0.4	14.4	0.6	19.5	0.8
Hb mmol/l **	3.8	0.3	8.9	0.4	12.1	0.5
Hct %	17.9	1.5	40.3	2.0	54.5	2.4
Hct l/l	0.179	0.015	0.403	0.020	0.545	0.024
MCV fl	79.5	5.0	91.0	5.0	97.5	5.0
MCH pg **	27.1	2.5	32.5	2.5	34.9	2.5
MCH fmol **	1.68	0.16	2.02	0.16	2.16	0.16
MCHC g/dl **	34.1	3.0	35.7	3.0	35.8	3.0
MCHC mmol/l **	21.2	1.9	22.2	1.9	22.2	1.9
RDW-CV	15.4	3.0	15.3	3.0	14.3	3.0
RDW-SD	43.7	6.0	49.7	6.0	49.8	8.0
Plts 10 ⁹ /l	71	20	234	40	468	60
MPV fl	8.4	3.0	8.9	3.0	8.6	3.0
LYMP %	58.5	12.0	28.4	8.0	12.8	7.0
MONO %	10.3	8.0	6.6	6.6	4.4	4.4
GRAN %	31.2	9.0	65.0	8.0	82.8	8.0
LYMP #	1.2	0.3	2.2	0.7	2.5	1.5
MONO #	0.2	0.2	0.5	0.5	0.8	0.8
GRAN #	0.7	0.2	5.0	0.7	16.0	1.7

** Mindray BC-3000 with CyMet Mindray (with CN) (old series s/n A59xxx)

Hb g/dl	5.8	0.4	13.4	0.6	18.2	0.8
Hb mmol/l	3.6	0.3	8.3	0.4	11.3	0.5
MCH pg	25.8	2.5	30.2	2.5	32.6	2.5
MCH fmol	1.60	0.16	1.88	0.16	2.02	0.16
MCHC g/dl	32.4	3.0	33.2	3.0	33.4	3.0
MCHC mmol/l	20.1	1.9	20.6	1.9	20.7	1.9

Avantor Performance Materials B.V.
P.O. Box 1 - 7400AA Deventer - The Netherlands
Tel.: +31 (0)570 687500

The devices as mentioned in this insert comply with the
In Vitro Diagnostic Medical Devices Directive 98/79/EC.



2°C

8°C