





The devices as mentioned in this insert comply with the  
In Vitro Diagnostic Medical Devices Directive 98/79/EC.

LOT : 218

🕒 : 2017-02-19

**Diatron\*\*:** Abacus, Abacus 3, Abacus Junior, Arcus; Erba: Elite 3

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	7.7	0.8	21.4	2.0
RBC 10 <sup>12</sup> /l	2.42	0.15	4.76	0.20	6.00	0.25
Hb g/dl	6.2	0.4	13.8	0.5	18.9	0.6
Hb mmol/l	3.8	0.3	8.6	0.4	11.7	0.5
Hct %	18.5	1.5	41.5	2.4	58.0	3.0
Hct l/l	0.185	0.015	0.415	0.024	0.580	0.030
MCV fl	76.6	5.0	87.2	5.0	96.7	5.0
MCH pg	25.6	2.4	29.0	2.8	31.5	2.8
MCH fmol	1.59	0.16	1.80	0.17	1.95	0.19
MCHC g/dl	33.4	3.0	33.3	3.0	32.5	3.0
MCHC mmol/l	20.7	2.0	20.7	2.0	20.2	2.0
RDW-SD fl	58.2	5.0	60.5	5.0	63.1	5.0
RDW-CV %	16.7	3.0	15.3	3.0	14.4	3.0
Plts 10 <sup>9</sup> /l	59	20	248	35	483	60
MPV fl	7.7	2.0	8.4	2.0	8.4	2.0
PCT %	0.05	0.02	0.21	0.06	0.41	0.10
PCT ml/l	0.45	0.20	2.08	0.60	4.06	1.00
PDW-SD fl	11.7	2.5	12.2	2.5	12.7	2.5
PDW-CV %	35.3	5.0	36.9	5.0	38.4	5.0
LYMP %	55.1	6.0	30.9	4.0	12.5	4.0
MONO %	9.1	9.1	5.9	5.9	6.4	6.4
GRAN %	35.7	6.0	63.3	5.0	81.1	6.0
LYMP #	1.1	0.3	2.4	0.4	2.7	0.9
MONO #	0.3	0.3	0.5	0.5	1.4	1.4
GRAN #	0.6	0.2	4.9	0.4	17.4	1.1

\*\* JT Baker reagents CyMet Abacus+Diluid Abacus=Use Reagent Code 2 in service settings!

**Medonic CA530, CA620**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	6.6	0.8	16.6	2.0
RBC 10 <sup>12</sup> /l	2.34	0.15	4.61	0.20	5.63	0.25
Hb g/dl	6.2	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.9	0.3	8.4	0.4	11.4	0.5
Hct %	16.5	1.5	36.8	2.4	49.8	3.0
Hct l/l	0.165	0.015	0.368	0.024	0.498	0.030
MCV fl	70.4	4.0	79.8	4.0	88.4	4.0
MCH pg	26.6	2.4	29.5	2.8	32.5	2.8
MCH fmol	1.65	0.16	1.83	0.17	2.02	0.19
MCHC g/dl	37.8	3.0	37.0	3.0	36.8	3.0
MCHC mmol/l	23.5	2.0	23.0	2.0	22.8	2.0
RDW %	15.1	3.0	17.4	3.0	16.9	3.0
Plts 10 <sup>9</sup> /l	58	20	234	35	444	60
MPV fl	7.2	2.0	7.8	2.0	7.7	2.0
LYMP %	65.9	9.0	37.7	6.0	18.8	4.0
MONO %	9.1	9.1	8.3	8.3	7.7	7.7
GRAN %	25.0	5.0	54.0	9.0	73.5	10.0
LYMP #	1.2	0.2	2.5	0.4	3.1	0.9
MONO #	0.2	0.2	0.5	0.5	1.3	1.3
GRAN #	0.5	0.2	3.6	0.4	12.2	1.1

**Inmesco HM1060; Rayto RT-7600; HTI Micros CC 20 Plus**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.4	7.7	0.8	20.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.54	0.20	5.51	0.25
Hb g/dl	6.1	0.4	13.9	0.5	19.4	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	12.1	0.5
Hct %	17.8	1.5	39.4	2.4	53.2	3.0
Hct l/l	0.178	0.015	0.394	0.024	0.532	0.030
MCV fl	77.1	4.0	84.9	4.0	94.1	4.0
MCH pg	25.9	2.4	30.6	2.8	35.2	2.8
MCH fmol	1.60	0.2	1.86	0.2	2.19	0.2
MCHC g/dl	34.4	3.0	35.3	3.0	36.5	3.0
MCHC mmol/l	21.4	2.0	21.4	2.0	22.7	2.0
RDW-CV %	12.7	3.0	11.9	3.0	11.0	3.0
RDW-SD fl	37.9	5.0	40.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	51	20	236	35	501	60
MPV fl	9.6	2.0	9.4	2.0	9.0	2.0
PCT %	0.05	0.02	0.22	0.06	0.45	0.10
PCT ml/l	0.49	0.20	2.21	0.60	4.51	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	53.9	8.0	28.8	5.0	11.9	4.0
MONO %	14.8	14.8	7.3	7.3	7.0	7.0
GRAN %	31.3	6.0	63.9	8.0	81.1	8.0
LYMP #	1.0	0.2	2.2	0.4	2.4	0.9
MONO #	0.3	0.3	0.6	0.6	1.4	1.4
GRAN #	0.6	0.2	4.9	0.8	16.5	1.1

**Avantor BeneSphera™ H32**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.4	7.7	0.8	20.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.54	0.20	5.51	0.25
Hb g/dl	6.1	0.4	13.9	0.5	19.4	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	12.1	0.5
Hct %	17.8	1.5	39.4	2.4	53.2	3.0
Hct l/l	0.178	0.015	0.394	0.024	0.532	0.030
MCV fl	77.1	4.0	84.9	4.0	94.1	4.0
MCH pg	25.9	2.4	30.6	2.8	35.2	2.8
MCH fmol	1.60	0.2	1.86	0.2	2.19	0.2
MCHC g/dl	34.4	3.0	35.3	3.0	36.5	3.0
MCHC mmol/l	21.4	2.0	21.4	2.0	22.7	2.0
RDW-CV %	12.7	3.0	11.9	3.0	11.0	3.0
RDW-SD fl	37.9	5.0	40.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	51	20	236	35	501	60
MPV fl	9.6	2.0	9.4	2.0	9.0	2.0
PCT %	0.05	0.02	0.22	0.06	0.45	0.10
PCT ml/l	0.49	0.20	2.21	0.60	4.51	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	53.9	8.0	28.8	5.0	11.9	4.0
MONO %	14.8	14.8	7.3	7.3	7.0	7.0
GRAN %	31.3	6.0	63.9	8.0	81.1	8.0
LYMP #	1.0	0.2	2.2	0.4	2.4	0.9
MONO #	0.3	0.3	0.6	0.6	1.4	1.4
GRAN #	0.6	0.2	4.9	0.8	16.5	1.1

**ERMA PCE-170, ERMAX-18, PCE-210**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	7.7	0.8	20.5	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.78	0.20	5.56	0.25
Hb g/dl	6.1	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	11.4	0.5
Hct %	17.8	1.5	41.4	2.4	55.6	3.0
Hct l/l	0.178	0.015	0.414	0.024	0.556	0.030
MCV fl	75.1	4.0	86.6	4.0	99.9	4.0
MCH pg	25.9	2.4	28.5	2.8	32.9	2.8
MCH fmol	1.61	0.16	1.77	0.17	2.04	0.19
MCHC g/dl	34.4	3.0	32.9	3.0	33.0	3.0
MCHC mmol/l	21.4	2.0	20.4	2.0	20.5	2.0
RDW %	17.0	3.0	16.1	3.0	15.0	3.0
Plts 10 <sup>9</sup> /l	69	20	242	35	455	60
MPV fl	7.7	2.0	8.1	2.0	8.0	2.0
PDW %	17.2	3.0	16.4	3.0	16.8	3.0
LYMP %	58.9	9.0	32.9	4.0	15.7	4.0
MONO %	8.0	8.0	6.4	6.4	5.0	5.0
GRAN %	33.1	5.0	60.7	9.0	79.3	10.0
LYMP #	1.1	0.2	2.5	0.4	3.2	0.9
MONO #	0.2	0.2	0.5	0.5	1.0	1.0
GRAN #	0.6	0.2	4.7	0.4	16.3	1.1

**Urit 2900 Plus**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	7.7	0.8	22.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.76	0.20	5.91	0.25
Hb g/dl	6.0	0.4	13.3	0.5	17.3	0.6
Hb mmol/l	3.7	0.3	8.3	0.4	10.7	0.5
Hct %	17.2	1.5	37.5	2.4	49.1	3.0
Hct l/l	0.172	0.015	0.375	0.024	0.491	0.030
MCV fl	72.6	4.0	78.9	4.0	83.1	4.0
MCH pg	25.4	2.4	27.9	2.8	29.3	2.8
MCH fmol	1.58	0.2	1.73	0.2	1.82	0.2
MCHC g/dl	35.0	3.0	35.4	3.0	35.3	3.0
MCHC mmol/l	21.7	2.0	22.0	2.0	21.9	2.0
RDW-CV %	15.5	3.0	15.1	3.0	14.7	3.0
RDW-SD fl	23.9	5.0	30.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	58	20	228	35	381	60
MPV fl	16.6	2.0	13.3	2.0	12.4	2.0
PCT %	0.10	0.02	0.30	0.06	0.47	0.10
PCT ml/l	0.96	0.20	3.03	0.60	4.72	1.00
PDW %	13.5	3.0	13.8	3.0	14.3	3.0
LYMP %	58.9	8.0	31.0	5.0	12.7	4.0
MONO %	8.0	8.0	4.1	4.1	4.3	4.3
GRAN %	33.1	6.0	64.9	8.0	83.0	8.0
LYMP #	1.1	0.2	2.4	0.4	2.8	0.9
MONO #	0.1	0.1	0.3	0.3	1.0	1.0
GRAN #	0.6	0.2	5.0	0.8	18.6	1.1

