



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **ASO-Plus** MANUFACTURE DATE: 05/24/2014  
 CATALOG#: ASO-100  
 LOT #: 55198 EXP. DATE: 09/2015

COMPONENTS:

1. ASO LATEX	LOT #:	220414B	EXP. DATE:	09/2016
2. POSITIVE CONTROL	LOT #:	050313W	EXP. DATE:	09/2015
3. NEGATIVE CONTROL	LOT #:	54296	EXP. DATE:	10/2016
4. GLYCINE BUFFER 20X	LOT #:	54588	EXP. DATE:	11/2016

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>CONTROLS</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
POSITIVE	050313W	09/2015	+	+
NEGATIVE	54296	10/2016	-	-

**TEST VALUES/RESULTS**

<u>CONTROLS</u>	<u>RESULT</u>	<u>PASS/FAIL</u>
POSITIVE (+)	+	PASS
NEGATIVE (-)	-	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

\_\_\_\_\_  
 Jacqueline Sanchez  
 Quality Control

DATE:

05/27/2014