



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **C-REACTIVE PROTEIN (CRP)** MANUFACTURE DATE: 06/04/2010
 CATALOG #: CRP-100
 LOT #: 26112 EXP. DATE: 04/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
CRP LATEX	CRP-A	C10100422B	04/2012
CRP POSITIVE CONTROL	CRP-B	190110R	07/2012
CRP NEGATIVE CONTROL	SNC	24975	09/2012
GLYCINE BUFFER (20X)	CRP-D	24181	07/2012

TOLERANCE LEVELS OF CONTROLS

<u>CONTROL</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u>
POSITIVE	190110R	07/2012	+	+
NEGATIVE	24975	09/2012	-	-

TEST VALUES/RESULTS

<u>CONTROL</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
POSITIVE	+	+	+	2/2
NEGATIVE	-	-	-	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 06/04/2010
 Quality Control