



CERTIFICATE OF ANALYSIS
FINISHED KIT QUALITY CONTROL

PRODUCT NAME: **RAPID PLASMA REAGIN (RPR)** MANUFACTURE DATE: 03/05/2010
CATALOG#: RPR-500G
LOT#: 24465 EXP. DATE: 01/2012

COMPONENTS: 1. RPR ANTIGEN LOT #: 717 EXP. DATE: 01/2012
2. RPR REACTIVE (1:4) CONTROL LOT #: 1546 EXP. DATE: 11/2012
3. RPR NON-REACTIVE CONTROL LOT #: 4345 EXP. DATE: 10/2012

TOLERANCE LEVELS OF SERUM CONTROLS

<u>CONTROL</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>
REACTIVE	1546	11/2012	REACTIVE
NON-REACTIVE	4345	10/2012	NON-REACTIVE

TEST VALUES/RESULTS

<u>CONTROL</u>	<u>1</u>	<u>2</u>	<u># PASS</u>
REACTIVE	REACTIVE	REACTIVE	2/2
NON-REACTIVE	NON-REACTIVE	NON-REACTIVE	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 03/05/2010