



**CERTIFICATE OF ANALYSIS**  
**FINISHED KIT QUALITY CONTROL**

PRODUCT NAME: **RAPID PLASMA REAGIN (RPR)**      MANUFACTURE DATE: 11/08/2010  
 CATALOG#: RPR-500  
 LOT#: 28802      EXP. DATE: 07/2012

COMPONENTS: 1. RPR ANTIGEN      LOT #: 721      EXP. DATE: 07/2012  
 2. RPR REACTIVE (1:4) CONTROL      LOT #: 1563      EXP. DATE: 03/2013  
 3. RPR NON-REACTIVE CONTROL      LOT #: 4360      EXP. DATE: 04/2013


**TOLERANCE LEVELS OF CONTROLS**

<u>CONTROL</u>	<u>LOT #</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>
REACTIVE	1563	03/2013	REACTIVE
NON-REACTIVE	4360	04/2013	NON-REACTIVE

**TEST VALUES/RESULTS**

<u>CONTROL</u>	<u>RESULT 1</u>	<u>RESULT 2</u>	<u># PASS</u>
REACTIVE	REACTIVE	REACTIVE	2/2
NON-REACTIVE	NON-REACTIVE	NON-REACTIVE	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	 Quality Control	DATE:	<u>11/08/2010</u>
-----------------	--	-------	-------------------