

TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-1G 100 Teco Finished Kit** MANUFACTURE DATE: 04/04/2011
CATALOG#: **URS-1G-100**
LOT #: 31767 EXP. DATE: 07/2013

COMPONENT: URS PAPER - GLUCOSE LOT #: 30882 EXP. DATE: 07/2013


TOLERANCE LEVELS OF CONTROLS

CONTROLS	LOT #	EXP. DATE	GLUCOSE RANGE
KOVA I	139036-4	01/2012	500 to $\geq$ 2000 mg/dl
KOVA II	139317-7	02/2012	Trace to 500 mg/dl
KOVA III	140400-1	03/2013	Negative (-)

TEST VALUES/RESULTS

CONTROLS	RESULT 1	RESULT 2	# PASS
LEVEL I	500	1000	2 / 2
LEVEL II	250	250	2 / 2
LEVEL III	-	-	2 / 2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	 Quality Control	DATE:	<u>04/04/2011</u>
-----------------	--	-------	-------------------